

**Declaration of Conformity \***  
In accordance with EN 45014:1998

We **Esco Micro Pte Ltd**  
of 21 Changi South Street 1  
Singapore 486777  
Tel: +65 65420833  
Fax: +65 65426920

hereby declare on our sole responsibility that the product:

**Model : Airstream Class II Biohazard Safety Cabinet**

has been designed and manufactured to conform with the following standards,  
when installed and used in accordance with the operating and maintenance manual:

- |                        |  |
|------------------------|--|
| • EN 50081 - 1 : 1992  | <b>Emissions</b>   |
| • EN 55022             | <b>Class A Conducted Emissions</b>                         |
| • EN 61000 - 3 - 2 / 3 | <b>Harmonic Current Emissions and Voltage Fluctuations</b> |
| • EN 55022             | <b>Class A Radiated Emissions</b>                          |
| • EN 50082 - 1 : 1997  | <b>Immunity</b>  |
| • EN 12469 : 2000      | <b>Microbiological Safety Cabinets</b>                     |
| • EN 61010 – 1 (1997)  | <b>Electrical Equipment Safety Requirements</b>            |

This declaration is valid also as declaration of conformity to the following directives:

- |              |   |
|--------------|---|
| • 73/23/EEC  | <b>Directive on low voltage</b>                   |
| • 89/336/EEC | <b>Directive on electromagnetic compatibility</b> |
| • 93/68/EEC  | <b>Directive on CE Marking</b>                    |
| • 98/37/EEC  | <b>Directive on machinery</b>                     |

More information may be obtained from Esco's authorised distributors located within the EU and a list of these parties and their necessary contact information is available on request from Esco.

A handwritten signature in black ink, appearing to read 'Lim Lay Yew', with a horizontal line underneath.

Lim Lay Yew  
Chief Executive Officer

\* Applicable only to 230V 50 Hz cabinets